



IN THE UNITED STATES PATENT AND TRADEMARK OFFICE

Applicant:	BLOMQUIST	Examiner:	Zoila E. Cabrera
Serial No.:	10/087,449	Group Art Unit:	2125
Filed:	FEBRUARY 28, 2002	Docket:	14485.141US01
Due Date:	July 2, 2004	Confirmation No.:	9068
Title:	PROGRAMMABLE MEDICAL INFUSION PUMP		

CERTIFICATE UNDER 37 CFR 1.8:

I hereby certify that this correspondence is being deposited with the United States Postal Service as first class mail, with sufficient postage, in an envelope addressed to: Commissioner for Patents, P.O. Box 1450, Alexandria, VA 22313-1450 on April 18, 2005.

By:   
Name: John C. Reich

RESPONSE

Commissioner for Patents  
P.O. Box 1450  
Alexandria, VA 22313-1450

Sir:

The applicants provide the following amendments and remarks in response to the office action mailed on October 18, 2004. No claims have been amended, and claims 1-25 are currently pending. In view of the following remarks, the applicant respectfully requests advancement of this application to allowance.

Claims 1-25 stand rejected as being anticipated by Estes et al. (US 2003/0114836). The applicant respectfully traverses this rejection.

**1. Claims 1-7 and 9-11**

The office action cites three new passages in maintaining this rejection. However, these passages still do not teach the claimed method and apparatus for generating a table on the computer, entering an operating parameter into at least one of the cells, and downloading the operating parameters into the pump.

For example, page 6, paragraph 60, lines 6-8 of Estes teaches, "The event markers can be logged into the pump and stored for later downloading or entered directly into the running software program." This passage teaches downloading from a pump to a PC